

Safe-T-Vue® 10 • Product No. 7201

Quality Assurance: Safe-T-Vue® 10 Non-Reversible Blood Bag Temperature Indicator Product No. 7201

PLEASE NOTE: ALL BOX OF 50 INFORMATION AND FOIL LID COPY FOR SAFE-T-VUE® 10 IS PRINTED IN BLACK.

Lot # **S10 032210**

Expiration Date: **2012-03**

Non - Chemistry

Mark \checkmark for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
\checkmark Copy	\checkmark Copy	\checkmark Copy	\checkmark Formed	\checkmark Acceptable
\checkmark Tack	\checkmark Tack	\checkmark Position	\checkmark Locked/Stable	Underfill
\checkmark Clean	\checkmark Color	\checkmark Clean Peel	Hinge	Overfill
\checkmark Box Damage	Tray	Disk	\checkmark Elasticity	Color Label
	\checkmark Edges	\checkmark Appearance		\checkmark Color
	\checkmark Color	\checkmark Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003

Data from a "DigiSense" Thermister Electronic Thermometer, using a cold (start temperature 1° to 3° C) bag of simulated blood (10% glycerol-water) with thermometer probe positioned in center of blood bag. Agitation of blood bag necessary for collection of accurate temperature-color change data. Download detailed protocol, from www.williamlabs.com, QAP 0300-01 REV 0 "Receiving Inspection – Temperature Indicators".

Temperature Change and Corresponding Safe-T-Vue® Change

Sample Set 1	DATA	Sample Set 2
1 to 4 °C	Starting Temperatures	1 to 4 °C

OBSERVATIONS

Temperature	Condition	
9.0 °C	Red Spots - Random	WHITE CENTER
9.7 °C	Red Filling In	WHITE CENTER
10.0 °C	Full Red Color	RED CENTER

Temperature	Condition	
9.0 °C	Red Spots – Random	WHITE CENTER
9.7 °C	Red Filling In	WHITE CENTER
10.0 °C	Full Red Color	RED CENTER

Temperature accuracy is +/-0.3°C.

Results of Testing: Red Spots - Random 9.0°C; Red Filling in White Center 9.7°C; Full Red Color at 10.0°C.



QUALITY ASSURANCE: APPROVED LOT # S10 032210 ; *Edward N. Sharpless* **March 26 2010**

Laboratory Director

Date