

Safe-T-Vue® 10 • Product No. 7201

Quality Assurance: Safe-T-Vue® 10 Non-Reversible Blood Bag Temperature Indicator Product No. 7201
 PLEASE NOTE: ALL BOX OF 50 INFORMATION AND FOIL LID COPY FOR SAFE-T-VUE® 10 IS PRINTED IN BLACK.

Lot # **S10 022411**
 Expiration Date: **2013-02**

Non - Chemistry

Mark √ for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
√ Copy	√ Copy	√ Copy	√ Formed	√ Acceptable
√ Tack	√ Tack	√ Position	√ Locked/Stable	Underfill
√ Clean	√ Color	√ Clean Peel	Hinge	Overfill
√ Box Damage	Tray	Disk	√ Elasticity	Color Label
	√ Edges	√ Appearance		√ Color
	√ Color	√ Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003

Data from a "DigiSense" Thermister Electronic Thermometer, using a cold (start temperature 1° to 3° C) bag of simulated blood (10% glycerol-water) with thermometer probe positioned in center of blood bag. Agitation of blood bag necessary for collection of accurate temperature-color change data. Download detailed protocol, from www.williamlabs.com, QAP 0300-01 REV 0 "Receiving Inspection – Temperature Indicators".

Temperature Change and Corresponding Safe-T-Vue® Change

Sample Set 1	DATA	Sample Set 2
1 to 4 °C	<i>Starting Temperatures</i>	1 to 4 °C


OBSERVATIONS

Temperature	Condition	
8.7 °C	Red Spots - Random	WHITE CENTER
9.4 °C	Red Filling In	WHITE CENTER
9.9 °C	Full Red Color	RED CENTER

Temperature	Condition	
8.8 °C	Red Spots – Random	WHITE CENTER
9.4 °C	Red Filling In	WHITE CENTER
9.9 °C	Full Red Color	RED CENTER

Temperature accuracy is +/-0.3°C.

Results of Testing: Red Spots - Random 8.8°C; Red Filling in White Center 9.4°C; Full Red Color at 9.9°C.



QUALITY ASSURANCE: APPROVED LOT # S10 022411 ; Edward N. Sharpless **March 11, 2011**

Laboratory Director

Date