

Safe-T-Vue[®] 6 • Product No. 7206

Quality Assurance: Safe-T-Vue[®] 6 Non-Reversible Blood Bag Temperature Indicator Product No. 7206
 PLEASE NOTE: ALL BOX OF 50 INFORMATION AND FOIL LID COPY FOR SAFE-T-VUE[®] 6 IS PRINTED IN PURPLE.

Lot # **S6 020911**

Expiration Date: February 2013

Non - Chemistry

Mark \checkmark for "Accepted" and X for "Not Accepted."

| Box | Adh. Label | Foil Lid | Close Function | Fill Level |
|-------------------------|--------------------|-------------------------|----------------------------|-------------------------|
| \checkmark Copy | \checkmark Copy | \checkmark Copy | \checkmark Formed | \checkmark Acceptable |
| \checkmark Tack | \checkmark Tack | \checkmark Position | \checkmark Locked/Stable | Underfill |
| \checkmark Clean | \checkmark Color | 0 Clean Peel | Hinge | Overfill |
| \checkmark Box Damage | Tray | Disk | | \checkmark Elasticity |
| | \checkmark Edges | \checkmark Appearance | | \checkmark Color |
| | \checkmark Color | \checkmark Fit | | |

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003

Data from a "DigiSense" Thermister Electronic Thermometer, using a cold (start temperature 1° to 3° C) bag of simulated blood (10% glycerol-water) with thermometer probe positioned in center of blood bag. Agitation of blood bag necessary for collection of accurate temperature-color change data. Download detailed protocol, from www.williamlabs.com, QCP/0200-5, "Receiving Inspection – Temperature Indicators". [Important: 7206 has a specific incoming inspection protocol.]

Temperature Change and Corresponding Safe-T-Vue[®] Change

| Sample Set 1 | DATA | Sample Set 2 |
|--------------|------------------------------|--------------|
| 1 to 3 °C | <i>Starting Temperatures</i> | 1 to 3 °C |

OBSERVATIONS

| Temperature | Color Change |
|-------------|------------------------------------|
| 5.1 °C | Red Spots - Random WHITE CENTER |
| 5.7 °C | Red Filling In WHITE CENTER |
| 6.1 °C | Full Red Color RED CENTER |

| Temperature | Color Change |
|-------------|------------------------------------|
| 5.2 °C | Red Spots - Random WHITE CENTER |
| 5.7 °C | Red Filling In WHITE CENTER |
| 6.1 °C | Full Red Color RED CENTER |

Temperature accuracy is +/-0.3°C.

Results: Red Spots - Random 5.2 °C; Red Filling in White Center 5.7 °C; Full Red Color at 6.1 °C.

