

TRANS-VUE™ 6 • Product No. 8206

Quality Assurance: TRANS-VUE 6 Non-Reversible Blood Bag Temperature Indicator Product No. 8206
PLEASE NOTE: ALL BOXES OF 50, ROUND COLOR-CODED LABEL, AND FOIL LID COPY FOR TRANS-VUE 6 IS PRINTED IN PURPLE.

Test performed on Random Sampling of Finished Product Lot T6 082610; Expiration Date: 2012-08

Non - Chemistry

Mark for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
<input checked="" type="checkbox"/> Copy	<input checked="" type="checkbox"/> Copy	<input checked="" type="checkbox"/> Copy	<input checked="" type="checkbox"/> Formed	<input checked="" type="checkbox"/> Acceptable
<input checked="" type="checkbox"/> Tack	<input checked="" type="checkbox"/> Tack	<input checked="" type="checkbox"/> Position	<input checked="" type="checkbox"/> Locked/Stable	<input type="checkbox"/> Underfill
<input checked="" type="checkbox"/> Clean	<input checked="" type="checkbox"/> Color	<input checked="" type="checkbox"/> Clean Peel	Hinge	<input type="checkbox"/> Overfill
<input checked="" type="checkbox"/> Box Damage	Tray	Disk	<input checked="" type="checkbox"/> Elasticity	Color Label
	<input checked="" type="checkbox"/> Edges	<input checked="" type="checkbox"/> Appearance		<input type="checkbox"/> Color
	<input checked="" type="checkbox"/> Color	<input checked="" type="checkbox"/> Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product.

Data collected from a DiGiSense® Model ThermoLogR Thermister Electronic Thermometer and compatible thermister direct contact 'button' probe. Using a ThermoSafe® Model BC-15 validated Blood Transport Cooler, 5 simulated blood bags (glycerol-water solution) and two cold packs, a TRANS-VUE 6 with probe taped to exposed media pillow is placed between and in contact with two of the simulated blood bags. For a detailed protocol request QCP/0200-5 from William Laboratories, Inc. 800.767.7643. [DiGiSense® is a registered trademark of Eutech Instruments; ThermoSafe® is a registered trademark of SCA Packaging North America.]

Temperature and Corresponding TRANS-VUE™ Color

Starting Temperature	°C
Trial 1	2.3
Trial 2	2.3
Trial 3	2.9

RESULTS	Full RED Color °C
Trial 1	5.8
Trial 2	6.0
Trial 3	6.0

Temperature accuracy is +/-0.3°C.

Results of Testing Lot # T6 082610 **Initial Validation;** Test Date: October 15, 2010

Full Red Color at 6.0° C.

