

Safe-T-Vue® 10 • Product No. 7201

Quality Assurance: Safe-T-Vue® 10 Non-Reversible Blood Bag Temperature Indicator Product No. 7201
PLEASE NOTE: ALL BOX OF 50 INFORMATION AND FOIL LID COPY FOR SAFE-T-VUE® 10 IS PRINTED IN BLACK.

Test performed on Random Sampling of Finished Product Lot # S10 030509
 Expiration Date: 2011-03

Non - Chemistry

Mark for “Accepted” and X for “Not Accepted.”

Box	Adh. Label	Foil Lid	Close Function	Fill Level
Copy	Copy	Copy	Formed	Acceptable
Tack	Tack	Position	Locked/Stable	Underfill
Clean	Color	Clean Peel	Hinge	Overfill
Box Damage	Tray	Disk	Elasticity	Color Label
	Edges	Appearance		Color
	Color	Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003

Performed by using a “DigiSense” Thermocouple Electronic Thermometer. Using a cold (2∞ to 6∞ C) bag of simulated blood (a glycerol-water-red dye solution) with the electric thermometer placed and sealed into the core of the blood bag to permit observation of temperature changes. For detailed protocol request QAP 0300-01 REV 0 Mar 01, 2005 [Incoming inspection of 7201 Safe-T-Vue® 10]

Temperature Change and Corresponding Safe-T-Vue® Change

Sample Set 1	DATA	Sample Set 2
1 to 4 ∞C	<i>Starting Temperatures</i>	1 to 4 ∞C

OBSERVATIONS

Temperature	Condition
9.1 ∞C	Red Spots - WHITE Random CENTER
9.8 ∞C	Red Filling In WHITE CENTER
10.0 ∞C	Full Red Color RED CENTER

Temperature	Condition
9.0 ∞C	Red Spots – WHITE Random CENTER
9.7 ∞C	Red Filling In WHITE CENTER
10.0 ∞C	Full Red Color RED CENTER

Temperature accuracy is +/-0.3°C.

Results of Testing Lot #S10 030509 Test Date: April 30, 2009

Red Spots - Random 9.1∞C; Red Filling in White Center 9.8∞C; Full Red Color at 10.0∞C.



